

Specification Sheet

Atomic Absorption Spectrophotometers

AA-7800 Series

The AA-7800 series Atomic Absorption Spectrophotometers can be configured to meet a wide range of applications. We can build the smallest dual atomizer system by adding a flame analysis system with high analytical performance and safety, and a furnace analysis system with simplicity and high performance. It is possible to analyze various samples at a wide range of concentrations using flexible instrument accessory configurations. In addition, software supports safe and secure data management.



Main Unit

Basics

Optics	Wavelength range	185.0 to 900.0 nm
	Monochromator	Aberration-corrected Czerny-Turner mounting, Number of grating grooves: 1800 lines / mm, Focal length: 300 mm
	Bandwidth	0.1, 0.2, 0.4, 0.7, 1.0, 2.0L nm (6-step automatic switching)
	Detector	Photomultiplier tube
	Optics	<ul style="list-style-type: none"> Optical double-beam Furnace: High-throughput single beam
	Background correction method	<ul style="list-style-type: none"> BGC-SR (high-speed self-reversal method) (185.0 to 900.0 nm) BGC-D2 (D₂ lamp method) (185.0 to 430.0 nm)
	Number of HC lamps	<ul style="list-style-type: none"> 8 lamps 2 lamps simultaneously lit (1 for measurement, 1 warming up for next measurement).
	Lamp mode	EMISSION, NON-BGC, BGC-D ₂ , BGC-SR
	Software requirements	Microsoft Windows 10 Pro (64bit)
	Parameter setting	Wizard method
Measurement mode	Flame continuous method, flame micro sampling method, furnace method, flame emission method	
Concentration conversion mode	<ul style="list-style-type: none"> Calibration curve method (order: 1st, 2nd, 3rd) Standard addition method, simple standard addition method (order: 1st) 	
Repeat analysis	Up to 20 repetitions. Mean value, standard deviation (SD) and coefficient of variation (RSD) display. Automatic exclusion of deviant values by setting SD and %RSD	
Baseline correction	Automatic correction of baseline drift by offset correction in peak height / peak area modes.	
Signal processing segment setting	Signal processing segments can be changed in peak height / peak area modes.	
Sensitivity correction	Automatic calibration curve correction function using sensitivity monitoring	
Data processing	Analog output	2 channels (atomic absorption/energy signal, background signal) Output range: 5.0, 2.5, 1.25, 0.625 Abs./V (each settable in 4 stages) Fixed at 1 V F.S. in EMISSION mode.
	Tabular data processing	Final concentration calculations based on sampled volume, dilution rate, fixed volume, and factor inputs
	Recall of parameters	Template functions available
	Procedure/result display	MRT (Measurement Results Table) worksheet
	Report generation	Summary report
	QA/QC	Select whether to continue or discontinue measurements based on results of evaluation on coefficient of correlation, %RSD, ICV-ICB, CCV-CCB, PB, LCS, SPK, PDS, and DUP.
	Re-analysis	<ul style="list-style-type: none"> Prescriptive and intelligent dilution options. Automatic dilution and re-analysis of unknown samples via autosampler (flame micro sampling method, furnace method)
	Digital recording	<ul style="list-style-type: none"> Management by login ID and password Control user access authority by user level Log record • Audit trail • Electronic signatures

Power requirements	Choose from 100, 120, 220, or 230 VAC, 230 VA, 50/60 Hz (Power is required separately for the personal computer.)
Dimensions and weight	<ul style="list-style-type: none"> AA-7800F/AAC: 702 W x 610 D x 702 H mm, 78 kg AA-7800F: 702 W x 610 D x 702 H mm, 75 kg AA-7800G: 702 W x 590 D x 520 H mm, 68 kg (Protruding parts and optional equipment are not included.)
Ambient temperature / humidity	10 to 35 °C, 20 to 80% (less than 70% when temperature is higher than 30 °C)

Required Software Operation Environment

The personal computer and monitor specifications are as follows.

Operating system	Microsoft Windows 10 Pro (64 bit)
CPU	Intel® Core i3-6100(3.70GHz) or higher is required.
RAM	4 GB or higher
Monitor	XGA (1280x1024 dots) or higher
Storage device	One CD-ROM drive (for installing a software) Requires 60 MB min. hard disk space for AA installation.
I/O port	One LAN port (for AA control) USB 2.0 port (for optional GFA-TV)
Peripheral device	Monitor, keyboard, mouse, printer

* Notes

- The PC, monitor and printer are not included in the standard configuration. Acquire them separately.
- The ASC-7800, ASK-7800 and GFA-7800 are optional accessories.
- A micro sampling kit is required for the flame micro sampling method.
- Microsoft Windows 10 Pro is a registered trademark of Microsoft Corp., U.S.A.
- The AA-7800 Series does not conform to IEC60601 — Medical Electrical Equipment.

Main Unit

Flame

Burner unit	Type	Air-cooled pre-mix type
	Burner head	Titanium 10 cm slot (5 cm titanium slot for N ₂ O-C ₂ H ₂ flame available as an option)
	Nebulizer	<ul style="list-style-type: none"> • Pt-Ir capillary • PTFE orifice • ceramic impact bead (capable of handling hydrofluoric acid)
	Chamber	Engineering plastics
	Positioning	AA-7800F <ul style="list-style-type: none"> • Lateral/vertical manual adjustment AA-7800F/AAC <ul style="list-style-type: none"> • Automatic flame/furnace switching by motor • Automatic search of optimum burner height
	Angle adjustment	0 to 90° (Angle adjustment is not possible if the optional GFA-7800 is installed on the AA-7800F/AAC.)
Gas control	Type	Air-C ₂ H ₂ flame, N ₂ O-C ₂ H ₂ flame (Hydrogen flame is not applicable)
	Flow rate control	<ul style="list-style-type: none"> • Automatic fuel gas flow rate setting (0.1 L/min step) • Manual support gas flow rate setting • Automatic search of optimum gas flow rate
	Safety measures	<ul style="list-style-type: none"> • Automatic flame extinction via flame vibration sensor • Acetylene regulator failure sensor • Automatic gas leak check • Automatic Air-N₂O switching as C₂H₂ flow rate increases • Flame monitor • Prevention of incorrect burner head use • Gas pressure monitor • Drain tank level monitor • Automatic flame extinction upon power outage or sudden power interruption • Internal fan stop sensor

Furnace (GFA-7800)

Heating temperature range	• Ambient to 3,000 °C
Heating control system	<ul style="list-style-type: none"> • Drying: Digital current control with automatic temperature calibration function • Ashing, Atomization: Digital temperature control via optical sensor
Setting heating conditions	<ul style="list-style-type: none"> • Maximum 20 stages • Heating mode: RAMP/STEP • Inner gas type: Dual automatic switching type • High-sensitivity mode setting • Enrichment in furnace: Maximum 20 times • Optimum temperature program search support function • Inner gas flow rate: 0 to 1.50 L/min
Safety measures	<ul style="list-style-type: none"> • Cooling water flow rate monitor • Gas pressure monitor • Overcurrent protection unit (double check by circuit protector and optical sensor) • Furnace block cooling check • Heat protection function
Positioning	AA-7800G / AA-7800F <ul style="list-style-type: none"> • Lateral/vertical manual adjustment AA-7800F/AAC <ul style="list-style-type: none"> • Automatic flame/furnace switching by motor
Power requirements	200, 220, 230, or 240 VAC ±5%, 7400 VA, 50/60 Hz
Dimensions and weight	230 W × 560 D × 507 H mm, 50 kg

Autosampler

Common Specifications

Control	RS-232C communication control from AA unit
Maximum reagent / sample positions	Reagents: 8 positions Samples: 60 positions (random access available with reagents or samples)
Rinse water bottle	2 L
Power requirements	Choose from 100, 120, 220, or 230 VAC, 50 VA, 50/60 Hz

Flame Continuous Method

Functions	Zero-point detection, auto diagnosis, auto rinse, random access
Sample volume	Sample vials: 15 mL, reagent vials: 53 mL
Nozzle rinse	Solvent rinse aspiration method
Dimensions and weight	Sampler: 340 W × 280 D × 270 H mm, 8 kg Controller (including stand): 230 W × 200 D × 320 H mm, 4.5 kg

Furnace Method / Flame Micro Sampling Method

Functions	Zero-point detection, auto diagnosis, auto rinse, auto mixing, random access
Sample volume	Sample vials: 2 mL, reagent vials: 20 mL
Nozzle rinse	Solvent rinse discharge method
Sampling functions	Dilution function, reagent addition function
Syringe	250 µL
Injection volume	2 to 90 µL
Repeatability	1%RSD (20 µL)
Carryover	Rinse port: Less than 0.0001 Mixing port: Less than 0.00001
Mixing port rinse	Solvent rinse discharge method, solvent discharge and rinse with next sample
Mixing function	Performed in mixing port. Maximum mixture volume: 0.6 mL
Maximum number of mixing reagents added	Up to 4 solutions Number of solutions possible for mixing <ul style="list-style-type: none"> • Calibration curve method: 5 solutions max. (sample + 4 types of reagents) • Standard addition method: 6 solutions max. (sample + standard solution + 4 types of reagents)
Auto dilution / re-analysis	For measurement result on unknown samples: <ul style="list-style-type: none"> • If extrapolation of calibration curve is possible: automatic calculation of dilution rate and dilution to bring concentration within calibration curve range • If extrapolation of calibration curve is not possible: dilution rate fixed at 10x
Dimensions and weight	Sampler (including sliding parts): 340 W × 280 D × 400 H mm, 11 kg Controller: 200 W × 110 D × 230 H mm (built into GFA-7800), 4 kg



Shimadzu Corporation

www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.

This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.

Company names, products/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation, its subsidiaries or its affiliates, whether or not they are used with trademark symbol "TM" or "®".

Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".

Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.